



**Quote Tabulation**  
Date: 11/07/2023

**ITB 2024-841, Medical & Surgical Equipment for Animal Shelter**

**Awarded Vendors**

Item	Description	Qty	McKesson Medical-Surgical Government Solutions		MWI Veterinary Supply Co.		Technology International, Inc.		Midwest Veterinary Supply, Inc		Medline Industries LP	
			Unit Price	Freight	Unit Price	Freight	Unit Price	Freight	Unit Price	Freight	Unit Price	Freight
1	Digital Xray System	1	\$58,814.65		\$43,888.89	\$2,439.25	\$9,800.00		\$48,944.00		X Ray Option 1 - \$53,950.00 add on to Option 1 \$1,950.0 Option2 - \$32,920.00 Imaging Table (Separate) Option 1 - \$13,585.00 Option 2 - \$10,725.00	
2	UV-C Air Purification System w/Light	5	No Bid		\$7,773.90		No Bid		\$8,300.00		No Bid	
3	Mounting Box	5	No Bid		No Bid		No Bid		\$2,665.00		No Bid	
4	UV-C Air Purification System w/o Light	5	No Bid		No Bid		No Bid		\$6,315.00		Portable Air Disinfection Device Option 1 - \$3,995.00 Accessories - \$975.00 & \$150.00 Option 2 - \$2,995.00 Accessories - \$675.00	
5	Ultrasound System	1	No Bid		\$6,000.00		\$3,180.00		\$5,019.00		Option 1 - \$10,335.00 Option 2 - \$6,300.00	
6	Dental Station	1	No Bid		\$4,511.37		\$5,900.00		\$5,497.00		Product - Price 5355C - \$3,810.17 0017-603-GRY-\$199.85 266040 - \$1,465.45 266064 - \$471.17 264563 - \$994.28 (Total \$6940.92)	
7	Veternary ECG Monitor	1	No Bid		\$5,593.00		\$1,000.00		\$5,993.00		\$3,684.80	
8	CO2 Leads	1	No Bid		No Bid		No Bid		\$1,991.00		No Bid	
Lead Time:			2-4 weeks after receipt of order		TBD		Aprox. 20 weeks after order and approval		2-6 weeks depending on item size		TBD	
Quote Expiration Date:			12/29/23		Remain firm through the award of the bid		30 days		Did not state		Valid until 9/28/23 Freight tax added at time of purchase	
<b>Total:</b>			<b>\$58,814.65</b>	<b>\$ -</b>	<b>\$67,767.16</b>	<b>\$2,439.25</b>	<b>\$19,880.00</b>	<b>\$ -</b>	<b>\$35,780.00</b>	<b>\$0.00</b>		<b>\$0.00</b>

Item 1 from Technology International Inc. was not awarded due to the Xray machine not meeting specification - Submitted Digital Portable x-Ray ZPXR-A10. Machine does not have table or monitor. This would require a separate screen to be purchased to view images.

Item from 4 Medline Industries LP submitted a Portable Air Disinfection Device and not what was requested, an Air Purification mounted in the ceiling is required.



## Quote Tabulation

Date: 11/07/2023

### ITB 2024-841, Medical & Surgical Equipment for Animal Shelter

Item 5 from Technology International Inc. did not meet specifications. Bid specified wireless, portable hand held unit, able to pair across multiple devices. Quoted item (Ultrasound System ZUS-A2) is a portable device larger than a hand held unit and is not wireless and has 2 transducer probe connectors.

Item 5 from Midwest Veterinary Supply Inc. submitted model that is not wireless and can only be view on Apple tablet or iphone with iOS8.0 or higher.

Item 6 from Medline Industries: Product 5355C is a only a cart with three handpieces. Product 0017-603-GRY is oxygen tubing. Product 266040 and 266064 are individual handpiece dental tools. Product 264563 is an individual hygiene handpiece. Dental station items submitted total \$6940.92.

Item 7 from Medline Industries LP did not meet specifications. Bid specified touchscreen, wireless networking, able to generate excel and pdf reports, remote access and measures ECG, SpO2, BP and temperature. Item submitted by Medline doesn't not meet all the above specks.