



St. Johns River Water Management District

Michael A. Register, P.E., Executive Director

4049 Reid Street • P.O. Box 1429 • Palatka, FL 32178-1429 • 386-329-4500 • www.sjrwmd.com

DATE: November 27, 2023
TO: Interested Firms
FROM: Breanna Pierce, Procurement Specialist
SUBJECT: Addendum 1 to RFP 39299 – Purchase of One (1) Total Phosphorus/Total Kjeldahl Nitrogen Continuous Flow Analyzer

As a result of questions received from potential respondents, please be advised of the following changes/information to the subject Request for Proposal:

1. The time and date for the submittal of responses remains the same: **2:00 p.m. Tuesday, December 5, 2023.**
2. **Question:** Is the requested system required to be two channels for the simultaneous detection of TKN and TP? The question is being asked as a one channel system is capable of TKN and TP analysis which meet specifications but does not perform the tests in a simultaneous manner.
Response: The instrument must be multi-channel to analyze for TP/TKN simultaneously.
3. **Question:** Are the shipping costs to be included with the total cost of the quoted system?
Yes. The Total Proposed Amount, including the TP/TKN system, on-site installation and training, and annual service and maintenance of TP-TKN system after first-year warrant, shall include delivery to the St. Johns River Water Management District, 4049 Reid Street, Palatka, Florida.

NOTE: Please acknowledge receipt of this Addendum in your submittal.

If you have any questions regarding this addendum, contact Breanna Pierce at (386) 643-1168 or via email at bpierce@sjrwmd.com.