

City of Chattanooga

Addendum #1

R175257 / B305199

Hazardous Materials Detection Equipment

(1) The due date for Bids has been changed. Sealed Bids should be delivered to the Purchasing Office prior to 2:00 PM on September 19, 2018.

(2) The following clarifications are issued in response to questions submitted by potential bidders:

(a) Item 3 - Quad mix calibration gas - 50% LEL, 50 ppm CO, 10 ppm H₂S, and 18% O₂ with a balance of nitrogen.

(b) Item 4 - 10 ppm Isobutylene gas - should be 100 ppm Isobutylene with a balance of air.

(c) Item 5 - Chlorine Calibration gas - 10 ppm chlorine with a balance of nitrogen.

(d) Item 6 - Ammonia Calibration gas - 50 ppm of Ammonia with a balance of nitrogen.

(e) Item 10 - Gas detection kit w/ incase calibration - Gas detection kit with incase calibration for RAE Systems MultiRAE. Kit includes standalone AutoRAE 2 calibration station, 1 wireless MultiRAE Lite (O₂/LEL/CO/H₂S/PID), sampling probe w/ coiled teflon tubing, 2 cylinders of calibration gas, 2 demand flow regulators and supporting accessories.

(3) Bidders should acknowledge receipt of this Addendum by signing a copy below and submitting the signed copy with their bid.

Receipt of Addendum #1 is hereby acknowledged. _____.