



## Jackson County Board of Commissioners

67 Athens Street  
Jefferson, Georgia 30549  
Phone: (706) 367-6309  
Fax: (706) 708-2505  
Email: myarbrough@jacksoncountygov.com

Date: May 17, 2022

To: Medical Equipment Vendors

From: Myrna Yarbrough, Purchasing Manager

**RE: Addendum 1 to Jackson County ITB 220015, Provide cardiac monitors, EMS, Jackson County GA.**

1. Please see answer to question below. All other information remains the same.
2. All other terms and conditions in ITB 220015 remain unchanged.
3. Any inquiries concerning Addendum 1 should be made to Myrna Yarbrough, Purchasing Manager, (706) 367-6309 or via email at [myarbrough@jacksoncountygov.com](mailto:myarbrough@jacksoncountygov.com)
4. Jackson County reserves the right to reject any and all proposals, to waive any technicalities or irregularities and to award the offer based upon the most responsive, responsible submission.
5. Bidders must acknowledge receipt of addenda, by either signing and attaching this copy, or by writing "Acknowledge Addenda #1" at bottom of the proposal.
6. Question: Other traditional monitor and defibrillator solutions on the market require battery calibrations and additional spare batteries per the manufacturer's recommendations. The Tempus Pro specifications do not require spare batteries; this is due to having a modular solution designed around battery technology for patient monitoring purposes. Furthermore, this 11+ hour battery does not require battery swaps in your daily workflow. Would your department be willing to accept these significant cost savings in a response to only providing one Tempus Pro battery per monitor? The Tempus LS defibrillator will have a spare battery built into the quote per recommended use and in response to this RFP. Answer: This is acceptable as the technology accomplishes the same thing as having spare batteries.