



CARDIAC MONITORS

INVITATION TO BID – REQUISITION NO. 2018-0216

ADDENDUM NO. ONE (1) – DATED JANUARY 25, 2018

The information given in this addendum is in addition to or supersedes conflicting information to the invitation to bid and is hereby made a part of the request.

Bid Specifications, Section C, “Minimum Device Requirements and Specifications”
is amended to add the following:

Item 1 - r. SPCO monitoring, to meet compliance requirements for NFPA 1584.

Please be sure to acknowledge receipt of this addendum on your bid form.